



CERTIFICATE



Certificate No. PCS 01095.01

SUCOFINDO INTERNATIONAL CERTIFICATION SERVICES

Graha Sucofindo B1 Floor - Jl. Raya Pasar Minggu Kav. 34 Jakarta 12780
Phone: +62-21-7983666 ext. 1040; Fax: +62-21-7987015 / 7987029; Email: cs.sics@sucofindo.co.id

Menganugerahkan kepada
Hereby grants to

PT. CHITOSE INTERNASIONAL Tbk.
Industri III No. 5 RT. 001 RW. 008, Utama, Cimahi Selatan, Cimahi
Jawa Barat, Indonesia

Sertifikat Kesesuaian Produk:

Product Conformity Certificate:

PRODUK <i>(Product)</i>	: 1. TEMPAT TIDUR PASIEN NON ELEKTRIK <i>(Non-Electric Medical Beds)</i> 2. TEMPAT TIDUR PASIEN ELEKTROMEDIK <i>(Electric Medical Beds)</i>
TIPE / MODEL <i>(Type / Model)</i>	: Lihat Lampiran <i>(See Attachment)</i>
SPESIFIKASI <i>(Specification)</i>	: Lihat Lampiran <i>(See attachment)</i>
MEREK <i>(Brand)</i>	: CHITOSE
PABRIK <i>(Factory/ Production Site)</i>	: PT. CHITOSE INTERNASIONAL Tbk. Industri III No. 5 RT. 001 RW. 008, Utama, Cimahi Selatan, Cimahi Jawa Barat, Indonesia

Telah memenuhi standar : **SNI IEC 60601-1:2014**
(Comply with standard)
Peralatan Elektromedik - Bagian 1: Persyaratan umum keselamatan dasar dan kinerja esensial

SNI IEC 60601-2-52:2014
Peralatan Elektromedik - Bagian 2-52: Persyaratan khusus keselamatan dasar dan kinerja esensial tempat tidur pasien

Regulasi/Skema Sertifikasi : **SKEMA SERTIFIKASI PRODUK SUCOFINDO**
(Regulation/ Certification Scheme)

Sertifikat ini berdasarkan skema yang terdiri dari: Audit Pabrik, Pengujian Dan Pengawasan Berkala (Tipe 5)

This certificate is based on the scheme consists of: Factory Audit, Testing and Surveillance (Type 5)

No. PCA: 1144/SERCO-III/PCS/2022



Darwin Abas

Ketua Badan Pelaksana
Chairman of Governing Board

This certificate is valid from March 01st, 2022 until February 28th, 2027

01.00



Sertifikat ini diberikan tunduk pada syarat-syarat pada "Product Certification Agreement"
This certificate is granted Subject to the provisions on "Product Certification Agreement"

Issued date: March 01st, 2022



4097549

SCI-2007A

Lampiran - Sertifikat No. PCS 01095.01
(Attachment to Certificate No.PCS 01095.01)
A. Tempat Tidur Pasien Non Elektrik (*Non-Electric Medical Beds*)

No.	Model	Spesifikasi (Specification)
1	CB-001	Manual Tanpa Crank
2	CB-001-D	Manual Tanpa Crank
3	CB-002	Manual Tanpa Crank
4	CB-002-D	Manual Tanpa Crank
5	CB-135-D-ST	Manual 1 Crank, 1 Cylinder
6	CB-3003-D-ST	Manual 3 Crank
7	CB-3011-D-ST	Manual 1 Crank, 1 Cylinder
8	CB-3012-D-ST	Manual 2 Crank, 1 Cylinder
9	CB-3012-D4-ST	Manual 2 Crank, 1 Cylinder
10	CB-003 (STRECHER)	Manual 2 Crank
11	CB-7011-D	Manual 1 Crank, 1 Cylinder
12	CB-7012-D	Manual 2 Crank, 1 Cylinder
13	CB-7003-D	Manual 3 Crank
14	OPTIMUS 1M	Manual 1 Crank
15	OPTIMUS 2M	Manual 2 Crank
16	OPTIMUS 3M	Manual 3 Crank

B. Tempat Tidur Pasien Elektromedik (*Electric Medical Beds*)

No.	Model	Spesifikasi (Specification)
1	CB-3300-T	230V, 50/60Hz, IPX4, Kelas II (<i>Class II</i>)
2	CB-0733-T	230V, 50/60Hz, IPX4, Kelas II (<i>Class II</i>)
3	OPTIMUS 3E	230V, 50/60Hz, IPX4, Kelas II (<i>Class II</i>)

