



# PT. GAYA INSTRUMENTASI NUMERIK

Jalan Gunung Rahayu No. 2 RT. 02 RW. 11 Kelurahan Pasirkaliki  
Kecamatan Cimahi Utara Kota Cimahi 40514  
Phone Office : (022) 2003491 Fax: (022) 82003637 Hotline: 08122224881  
www.ginumerik.com E-mail : info@ginumerik.com



## SERTIFIKAT KALIBRASI

Calibration Certificate

Halaman : 1 dari 2  
Page of

### NOMOR SERTIFIKAT

Certified Number

: 2170.G-Sert/05/2023

No. Order

23 G 002 27

### Tanggal Kalibrasi

Calibration Date

: 24 Mei 2023

### IDENTITAS ALAT

Instrument Identity

Nama

Name

: Vernier Calliper

Merk / Pabrik

Merk / Manufacture

: Mitutoyo

Tipe / No seri

Type / Serial Number

: CD-12"C / 1108097

Kapasitas / Resolusi

Capacity / Resolution

: 0 - 300 mm / 0.01 mm

### IDENTITAS PEMILIK

Owner Identity

Nama

Name

: PT. CHITOSE INTERNATIONAL TBK

Alamat

Address

: Jl. Industri III No. 5 Utama, RT 001 RW 008, Cimahi, Jawa Barat 40533

### KONDISI RUANG

Environmental Condition

: T = ( 20.4 ± 0.5 ) °C RH = ( 56 ± 2.7 ) %

### IDENTITAS STANDAR

Standard Identity

Nama

Name

: Calliper Checker

No Seri

Serial Number

: 230128

### METODE KALIBRASI

Calibration Methode

: JIS B 7507 : 2016

Standar tersebut diatas telah tertelusur ke Sistem Satuan Internasional ( SI ) Melalui LK-001-IDN  
Ketidakpastian kalibrasi ini merupakan ketidakpastian bentangan yang didapat dari sumber - sumber  
kesalahan tipe A dan tipe B sesuai dengan JCGM 100:2008, yang dinyatakan pada tingkat kepercayaan  
95 % dengan faktor cakupan k = 2

Instrument Stated above has been traceable to System International Standard ( SI ), through LK-001-IDN  
This calibration uncertainty is Expanded Uncertainty wich referred to type A and B source of error according to JCGM 100:2008  
which is expressed as a 95 % confidence level by the coverage factor k = 2

Di terbitkan tanggal, 26 Mei 2023

Date of Issue

Disetujui Oleh  
Approved By

PT. GAYA INSTRUMENTASI NUMERIK  
Dycki Bachrul Rizal  
Approval Signatory

26 Mei 2023

Di verifikasi Oleh, mengetahui.

Yulan Septian  
Jr. Staf QC

shanty M.  
Manager

Tidak dibonarkan mengutip/memperbanyak dan/atau mempublikasikan sebagian isi sertifikat ini tanpa ijin PT. Gaya Instrumentasi Numerik  
Sertifikat ini sah bila telah dibubuhi cap PT. Gaya Instrumentasi Numerik dan ditandatangani oleh pejabat yang berwenang

Prohibited quote/ reproduce and/ or publish some of the content of this certificate without the written permission of PT. Gaya Instrumentasi Numerik  
This certificate valid if it has been stamped PT. Gaya Instrumentasi Numerik