



ISO 9001:2015 - 2017

PT. GLOBAL QUALITY INDONESIA

CALIBRATION OF INDUSTRIAL INSTRUMENT AND MEDICAL INSTRUMENT TESTING, TRAINING, QUALITY CONSULTANT,
PROFICIENCY TESTING, INSPECTION, CONFORMITY TESTING, CERTIFICATE PERSONAL, MAINTENANCE & REPAIR



Kemeterian Akreditasi Nasional

LK-057-IDN

CALIBRATION CERTIFICATE

Certificate Number : 07691/GQI-Ser/05/21

Page : 1 of 1

Order Number : 2104-181-2

Received Date : 29 April 2021

Equipment Name : Roll Meter

Manufacturer : Tajima

Model/Type : Hsok - 16

Serial Number : RM - 2

Capacity/Graduation : 3600 mm / 1 mm

Technician ID : T1

Typewriter ID : T8

Owner : PT. Chilose International Tbk

Address : J. Industri III No 5 RT 001/RW 008, Utama, Cimahi, Jawa Barat

Calibration Location : PT. Global Quality Indonesia

Calibration Date : 30 April 2021

Calibration Method : IK-D-11 ref. JIS B 7512 - 2016

Environmental Condition : Tbefore = 21.1 °C Rbefore = 53 %

Tafter = 21.1 °C Rafter = 52 %

Calibration Report

Equipment Indication (mm)	Standard Indication (mm)	Correction (mm)
50	50.000	0.000
100	100.000	0.000
200	200.010	0.010
300	300.010	0.010
400	400.015	0.015
500	500.015	0.015
1000	1000.020	0.020
1500	1500.025	0.025
2000	2000.030	0.030
2500	2500.035	0.035
3000	3000.035	0.035
3600	3600.040	0.040
Uncertainty $U_{95\%} = \pm 0.8 \text{ mm}$		

The Uncertainty is taken at a Confidence Level 95 % and Coverage Factor (k) = 2

Standard used

Name

Mark/Type

Serial Number

Traceable to SI through

LVDT

Size/Size

EDLD1364

LK-051-IDN

Issuance Date : 03 May 2021



GLOBAL QUALITY INDONESIA

Dir. Rudy Wardani

Director

End of Certificate

1. This calibration result is valid only from the equipment labelled.
2. It is not permitted to reproduce this certificate without permission from PT. Global Quality Indonesia.
3. The original calibration certificate of PT. Global Quality Indonesia uses a barcode mark.

Attention

FR.15.01



Komplek Kopo Mas Regency
Blok N No. 7C Bandung 40227



022-5436533 - 5436775
022-54425368



calibration@globalqualityindonesia.id
customer@globalqualityindonesia.id



076910-2117
Calibration Certificate No. 0000 0000



www.globalqualityindonesia.id

certificate