

# PT. GLOBAL QUALITY INDONESIA

CALIBRATION OF INDUSTRIAL INSTRUMENT AND MEDICAL INSTRUMENT, TESTING, TRAINING, QUALITY CONSULTANT,  
PROFICIENCY TESTING, INSPECTION, CONFORMITY TESTING, CERTIFICATE PERSONAL, MAINTENANCE & REPAIR



LK-057-IDN  
PUP-018-IDN  
LP-1707-IDN

## CALIBRATION CERTIFICATE

Certificate Number : 11507/GQI-Sert/10/23  
Page : 1 of 2

Order Number : 202310V0ZM-2	Owner : PT. Chitose Internasional Tbk
Received Date : 09 October 2023	Address : Jl. Industri III No.5, Leuwigajah Cimahi, Utama, Kec. Cimahi Selatan, Kota Cimahi, Jawa Barat 40533
Equipment Name : Vernier Caliper	Calibration Location : PT. Global Quality Indonesia
Manufacturer : Mitutoyo	Calibration Date : 11 October 2023
Model/Type : 530-502 / Analog	Calibration Method : IK-D-07 ref. JIS B 7507 : 2016
Serial Number : 16005738	Environmental Condition : Tbefore = 20,1 °C RHbefore = 50 %
Capacity/Resolution : 1000 mm / 0,05 mm	Tafter = 20,3 °C RHafter = 51 %
Technician ID : PTT	
Typewriter ID : PMI	

Calibration Report :

### Outside

Standard Indication (mm)	Instrument Indication (mm)	Correction (mm)
0	0,00	0,00
50	50,00	0,00
75	75,00	0,00
100	100,00	0,00
125	125,00	0,00
150	150,00	0,00
175	175,00	0,00
200	200,00	0,00
250	250,00	0,00
300	300,00	0,00
400	400,00	0,00
500	500,00	0,00
600	600,00	0,00
700	700,00	0,00
800	800,00	0,00
900	900,00	0,00
1000	1000,00	0,00
Uncertainty $U_{95\%} = \pm 0,03 \text{ mm}$		


The Uncertainty is Taken at a Confidence Level 95 % and Coverage Factor ( k ) = 2

### Standard Used :

Name :	Merk/Type :	Serial Number :	Traceable to SI Through :
Gauge Block	Mitutoyo/Grade 0	31121	LK-045-IDN
Long Gauge Block	Mitutoyo/Grade 1	9905314	SNSU BSN

Issuance Date : 13 October 2023

  
**GLOBAL QUALITY INDONESIA**  
Fidi Rudy Hamid  
Director

14 November 2023  
Diverifikasi oleh  
  
Yulan Septian  
Jr. staff QC

Mengetahui  
  
shanty. M.  
Manager QC.



FR.15.01

Attention : 1. This calibration result is valid only from the equipment calibrated  
2. It is not permitted to reproduce this certificate without permission from PT. Global Quality Indonesia

certificate

## CALIBRATION CERTIFICATE

Certificate Number : 11507/GQI-Sert/10/23  
 Page : 2 of 2  
 Order Number : 202310V0ZM-2  
 Received Date : 09 October 2023  
 Equipment Name : Vernier Caliper  
 Manufacturer : Mitutoyo  
 Model/Type : 530-502 / Analog  
 Serial Number : 16005738  
 Capacity/Resolution : 1000 mm / 0,05 mm

Calibration Report :

**Inside**

Standard Indication (mm)	Instrument Indication (mm)	Correction (mm)
0	0,00	0,00
50	50,00	0,00
75	75,00	0,00
100	100,00	0,00
125	125,00	0,00
150	150,00	0,00
175	175,00	0,00
200	200,00	0,00
250	250,00	0,00
300	300,00	0,00
400	400,00	0,00
500	500,00	0,00
600	600,00	0,00
700	700,00	0,00
800	800,00	0,00
900	900,00	0,00
1000	1000,00	0,00
Uncertainty $U_{95\%} = \pm 0,03$ mm		

The Uncertainty is Taken at a Confidence Level 95 % and Coverage Factor ( k ) = 2

— End of Certificate —




GLOBAL QUALITY INDONESIA