

PT. GLOBAL QUALITY INDONESIA

CALIBRATION OF INDUSTRIAL INSTRUMENT AND MEDICAL INSTRUMENT, TESTING, TRAINING, QUALITY CONSULTANT,
PROFICIENCY TESTING, INSPECTION, CONFORMITY TESTING, CERTIFICATE PERSONAL, MAINTENANCE & REPAIR



CALIBRATION CERTIFICATE

Certificate Number : 11608/GQI-Sert/10/23
Page : 1 of 1

Order Number : 202310V0ZM-1
Received Date : 09 October 2023
Equipment Name : Moisture Meter
Manufacturer : Sanfix
Model/Type : GM610
Serial Number : 2644243
Capacity : 2~70 %
Resolution : 0,5 %
Technician ID : GJ
Typewriter ID : PMI

Owner : PT. Chitose Internasional Tbk
Address : Jl. Industri III No.5, Leuwigajah Cimahi,
Utama, Kec. Cimahi Selatan, Kota
Cimahi, Jawa Barat 40533
Calibration Location : PT. Global Quality Indonesia
Calibration Date : 13 October 2023
Calibration Method : IK-I-40
Environmental Condition : Tinitial= 24,4 °C RHinitial= 55 %
Tfinal= 24,2 °C RHfinal= 53 %

Calibration Report :

Standard Indication		Instrument Indication	Correction
(MΩ)	(%)	(%)	(%)
470	12	11,5	0,5
210	13	12,0	1,0
100	14	13,5	0,5
52	15	14,0	1,0
28	16	15,0	1,0
16	17	16,5	0,5
10	18	17,0	1,0
6,8	19	18,0	1,0
4,4	20	19,5	0,5
2,8	21	20,0	1,0
1,9	22	21,5	0,5

Uncertainty $U_{95\%} = \pm 0,58 \%$

The Uncertainty is Taken at a Covidence Level 95 % and Coverage Factor (k) = 2

Standard Used :

Name of Calibrator : Merk/Type : Serial Number : Traceable to SI Through :
Decade Resistor Aerotech / 1G Series 151204 LK-032-IDN

Issuance Date : 17 October 2023



14 November 2023.
Diverifikasi Oleh.
[Signature]
Yulan Septian
Jr. Staff QC
— End of Certificate —

Mengetahui
[Signature]
Shanty M
Manajer QC.



certificate

FR.15.01

Attention : 1. This calibration result is valid only from the equipment calibrated
2. It is not permitted to reproduce this certificate without permission from PT. Global Quality Indonesia